




Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS) [*]	Cylinder Power (DC)	Axis ^o	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t [†]	UV Blocking [‡] & Class	Visibility Tint	FDA Group
 MyDay [®] daily disposable	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +8.00 (0.50 steps)	---	---	---	Asphere	Aberration Neutralizing System	Daily wear; 1 day replacement	8.4	14.2	Aquaform Technology	stenfilcon A	54	100	Class 2	Yes	5B (SiHy)
 MyDay [®] toric	-6.00 to Plano (0.25 steps) -10.00 to -6.50 (0.50 steps) +0.25 to +6.00 (0.25 steps)	-0.75, -1.25, -1.75 -2.25	10 to 180 (10 steps) 10-20, 90, 160-180	---	Toric	Optimised Toric Lens Geometry	Daily wear; 1 day replacement	8.6	14.5	Aquaform Technology	stenfilcon A	54	80	Class 2	Yes	5B (SiHy)
 MyDay [®] multifocal	-10.00 to -6.50 (0.50 steps) +0.50 to +6.00 (0.50 steps)	-2.25	10-20, 90, 160-180	Low (+0.75 to +1.25 spectacle add) Med (+1.50 to +1.75 spectacle add) High (+2.00 to +2.50 spectacle add)	Multifocal	Binocular Progressive System	Daily wear; 1 day replacement	8.4	14.2	Aquaform Technology	stenfilcon A	54	100	Class 2	Yes	5B (SiHy)

* Plano lens availability for sphere product can vary by market or customer.





† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂) / (ml x mmHg)]

‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS) [*]	Cylinder Power (DC)	Axis ^o	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t [†]	UV Blocking [‡] & Class	Visibility Tint	FDA Group
 Proclear® 1 day	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +8.00 (0.50 steps)	---	---	---	Asphere	Aberration Neutralizing System	Daily wear; 1 day replacement	8.7	14.2	PC Technology (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 Proclear® 1 day multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to Plano (0.25 steps) +0.25 to +6.00 (0.25 steps)	---	---	Single power profile	Multifocal	---	Daily wear; 1 day replacement	8.7	14.2	PC Technology (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 MiSight® 1 day	-10.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps)	---	---	Treatment zones offer +2.00D of myopic defocus	Dual-focus	ActivControl Technology	Daily wear; 1 day replacement	8.7	14.2	PC Technology (phosphoryl choline)	omafilcon A	60	28	No	Yes	2
 clariti® 1 day multifocal	-6.00 to Plano (0.25 steps) +0.25 to +5.00 (0.25 steps)	---	---	Low: Up to +2.25 High: +2.50 to +3.00	Multifocal	---	Daily wear; 1 day replacement	8.6	14.1	WetLoc Technology	somofilcon A	56	86	Class 2	No	5B (SiHy)

* Plano lens availability for sphere product can vary by market or customer.





† (@-3.00DS) x 10⁻⁸ [(cm/sec) x (ml O₂)/(ml x mmHg)]

‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS) [†]	Cylinder Power (DC)	Axis [°]	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t [‡]	UV Blocking [†] & Class	Visibility Tint	FDA Group
 clariti® 1 day sphere	-10.00 to -6.50 (0.50 steps)	— — —	— — —	— — —	Asphere	— — —	Daily wear; 1 day replacement	8.6	14.1	WetLoc Technology	somofilcon A	56	86	Class 2	No	5B (SiHy)
	-6.00 to -0.50 (0.25 steps)															
	+0.50 to +6.00 (0.25 steps)															
	+6.50 to +8.00 (0.50 steps)															
 clariti® 1 day toric	-9.00 to -6.50 (0.50 steps)	-0.75 -1.25 -1.75	10, 20, 70, 80, 90, 100, 110, 160, 170, 180	— — —	Toric	Smooth-gradient ballast toric design	Daily wear; 1 day replacement	8.6	14.3	WetLoc Technology	somofilcon A	56	57	Class 2	No	5B (SiHy)
	-6.00DS to Plano (0.25 steps)	-2.25	10, 20, 90, 160, 170, 180													
	+0.25 to +4.00 (0.25 steps)	-0.75 -1.25 -1.75	10 to 180 (10 steps)													
		-0.75 -1.25 -1.75	10, 20, 70, 80, 90, 100, 110, 160, 170, 180													
 Biomedics® 1 day Extra	-10.00 to -6.50 (0.50 steps)	— — —	— — —	— — —	Sphere	— — —	Daily wear; 1 day replacement	8.6	14.2	— — —	ocufilcon D	55	26	No	Yes	4
	-6.00 to -0.25 (0.25 steps)							8.8								
	+0.25 to +5.00 (0.25 steps)															
 Biomedics® 1 day Extra toric	-10.00 to -7.50 (0.50 steps)	-0.75, -1.25 -1.75	90, 180	— — —	Toric	Optimised Toric Lens Geometry	Daily wear; 1 day replacement	8.7	14.5	— — —	ocufilcon D	55	18	No	Yes	4
	-7.00 to -6.50 (0.50 steps)		20, 90, 160, 180													
	-6.00 to Plano (0.25 steps)															

[†] Plano lens availability for sphere product can vary by market or customer.

[‡] (@-3.00DS) x 10⁻⁹[(cm/sec) x (ml O₂)/(ml x mmHg)]

^{††} UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional. Biomedics®, clariti®, CooperVision®, MyDay®, MiSight® and Proclear® are registered trademarks of The Cooper Companies, Inc. and its subsidiaries.

Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking‡ & Class	Visibility Tint	FDA Group
 Biofinity®	-12.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps) NO PLANO	---	---	---	Asphere	Aberration Neutralizing System	Daily wear; 30 days replacement Extended wear 7 days/6 nights; 30 day replacement Continuous wear 30 days/nights; 30 day replacement (NOT IN EU/CE mark countries) - to be removed globally from 2024	8.6	14.0	Aquaform Technology	comfilcon A	48	171	No	Yes	5C (SiHy)
 Biofinity® XR	-20.00 to -12.50 (0.50 steps) +8.50 to +15.00 (0.50 steps)	---	---	---	Asphere	Aberration Neutralizing System	As for Biofinity sphere	8.6	14.0	Aquaform Technology	comfilcon A	48	171	No	Yes	5C (SiHy)
 Biofinity® toric	-10.00 to -6.50 (0.50 steps) -6.00 to +6.00 (0.25 steps) +6.50 to +8.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	---	Toric	Optimized Toric Lens Geometry	As for Biofinity sphere	8.7	14.5	Aquaform Technology	comfilcon A	48	116	No	Yes	5C (SiHy)

* Plano lens availability for sphere product can vary by market or customer.

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂) / (ml x mmHg)]

‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS)*	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking‡ & Class	Visibility Tint	FDA Group
 Biofinity® XR Toric	-20.00 to -10.50 (0.50 steps) +8.50 to +20.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	5 to 180 (5 steps)	— —	Toric	Optimized Toric Lens Geometry	As for Biofinity sphere	8.7	14.5	Aquaform Technology	comfilcon A	48	116	No	Yes	5C (SiHy)
	-20.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +20.00 (0.50 steps)	-2.75 -3.25 -3.75 -4.25 -4.75 -5.25 -5.75														
 Biofinity® multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to plano (0.25 steps) +0.25 to +6.00 (0.25 steps)	— —	— —	+1.00 +1.50 +2.00 +2.50	Multifocal D lens N Lens	Balanced Progressive Technology	As for Biofinity sphere	8.6	14.0	Aquaform Technology	comfilcon A	48	142 (-3.00, N lens, +1.00 Add)	No	Yes	5C (SiHy)
 Biofinity® toric multifocal	-10.00 to -6.50 (0.50 steps) -6.00 to +6.00 (0.25 steps) +6.50 to +10.00 (0.50 steps)	-0.75 to -5.75 (0.50 steps)	5 to 180 (5 steps)	+1.00 +1.50 +2.00 +2.50	Toric Multifocal D Lens N Lens	Optimized Toric Lens Geometry Balanced Progressive Technology	As for Biofinity sphere	8.7	14.5	Aquaform Technology	comfilcon A	48	116	No	Yes	5C (SiHy)

* Plano lens availability for sphere product can vary by market or customer.




† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mmHg)]

‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
<https://coopervision.com.hk>



Product	Sphere Power (DS) [†]	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Design Technology	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material Technology	Material USAN	Water content (%)	Oxygen transmissibility Dk/t [‡]	UV Blocking* & Class	Visibility Tint	FDA Group
 Proclear® multifocal toric	-20.00 to -6.50 (0.50 steps) -6.00 to Plano (0.25 steps) +0.25 to +6.00 (0.25 steps) +6.50 to +20.00D (0.50 steps)	-0.75 to -5.75 (0.50 steps)	5 to 180 (5 steps)	+1.00 to +4.00 (0.50 steps)	Toric Multifocal D lens N lens	— —	Daily wear; 30 days replacement 2 weekly replacement US	8.4 8.8	14.4	PC Technology (phosphoryl choline)	omafilcon B	62	From: 14.5 (-3.00D/+4.00D, N Type) To: 17.7 (-3.00D/+4.00D, D Type)	No	Yes	2
 Biomedics® Now (Sphere)	-10.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps)	— —	— —	— —	Asphere	Aberration Neutralizing System	Daily wear; 30 days replacement EU/CE mark markets (ANZ): Extended wear 7 days/6 nights; 30 day replacement (NOT IN EU/ANZ after 26 May 2024)	8.6	14.2	— —	ocufilcon D	55	26	Class 2	Yes	4
 Biomedics® toric	-9.00 to -6.50 (0.50 steps) -6.00 to -0.25 (0.25 steps) +0.25 to +5.00 (0.25 steps) +5.50 to +6.00 (0.50 steps)	-0.75 -1.25 -1.75 -2.25	10 to 180 (10 steps)	— —	Toric	— —	Daily wear; 30 days replacement EU/CE mark markets (ANZ): Extended wear 7 days/6 nights; 30 day replacement (NOT IN EU/ANZ after 26 May 2024)	8.7	14.5	— —	ocufilcon D	55	18	Class 2	Yes	4

* Plano lens availability for sphere product can vary by market or customer.

[†] (@-3.00DS) x 10⁻³ [(cm/sec) x (ml O₂)/(ml x mmHg)]

[‡] UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Biofinity®, Biomedics®, clariti®, CooperVision® and Proclear® are registered trademarks of The Cooper Companies, Inc. and its subsidiaries.

©2023 CooperVision SA10395 11/2023