

## Biofinity® multifocal lens fitting guidelines

# A simplified fitting philosophy

Our new fitting process is based on eye care professionals' real-world experiences.

- Fitting lower ADD powers is now simpler than ever, by using the same D lens design for both eyes
- Fitting higher ADD powers continues to be flexible, giving you more options for exceptional vision performance



### Initial visit

**Step 1** Start with a new refraction and verification of eye dominance (fogging technique).

**Step 2** Select the distance prescription based on spherical equivalent corrected for the vertex distance. Choose D or N lens design based on needed ADD power:

ADD	Dominant eye	Non-Dominant eye
+1.00	<b>D</b>	<b>D</b>
+1.50	<b>D</b>	<b>D</b>
+2.00	<b>D</b>	<b>N</b>
+2.50	<b>D</b>	<b>N</b>

Visual acuity expectations when using D and N lens combination

Lens	Distance	Near
Binocularly	20/20	20/20
D Lens	20/20	20/40 or better
N Lens	20/40 or better	20/20

**Step 3** Although lens will equilibrate quickly, allow patients to adapt to lenses for a minimum of 15 minutes before assessing vision. If binocular vision is unacceptable, perform a monocular over refraction, using hand-held trial lenses, to determine which eye needs improvement. To improve distance vision add +/-0.25D (up to +/-0.50D) to the eye that needs improvement. To improve near vision add +/-0.25D (up to +/-0.50D) to the eye that needs improvement.

**Clinical Tips** Prescribe maximum plus power for distance vision (Do not over minus)

Test patient's near function vision with their cell phone

Choose the lower ADD power when possible; not necessary to overprescribe the ADD power

Check visual acuity with room lights on

# A unique multifocal lens for unique eyes

## Balanced Progressive™ Technology

- Optimized for exceptional vision at all distances: near, intermediate, and far
- Allows for personalized fitting for each wearer and each eye
- Streamlined fitting process helps ensure success for presbyopic patients

For additional fitting tips, tutorials, and more information on Biofinity multifocal, visit [www.coopervision.com/multifocal](http://www.coopervision.com/multifocal)

The eye care professional retains the independent clinical judgment on how to fit and prescribe lenses.



## Follow-up visit one week later

If patient requires further enhancement to distance or near visual acuity.

**Step 1** Evaluate binocular visual acuity.

**Step 2** Check monocular visual acuity.

**Step 3** Perform over refraction using hand-held trial lenses (avoid using a phoropter).

FIRST OPTION: To improve either distance or near vision, modify distance vision by +/- 0.25D in the eye that needs improvement.

SECOND OPTION: To improve near vision add +0.50D to the ADD power of the eye that needs improvement.

## Product specifications

### Biofinity® multifocal

Base Curve	8.6 mm
Diameter	14.0 mm
Sphere Power	+6.00D to -8.00D (0.50D after -6.00D)
ADD Power	+1.00, +1.50, +2.00, +2.50
Lens Design	D Lens, N Lens
Material	comfilcon A
Water content	48%
Dk	128
Wearing schedule	Daily Wear or Extended Wear up to 6 nights/7 days



CooperVision™

[www.coopervision.com](http://www.coopervision.com)  
800 341 2020